PrymaLab
PRYMALAB ANALYTICAL · IN-HOUSE LABORATORY
3705 Music Branch Rd, Ashland, KY 41102 · +1 (606) 465-9063 · prymalab.net

CERTIFICATE NO.
PL-26-00024
VERIFICATION KEY
GFEZEMLN
Scan or enter both at prymalab.net/testing/verify

CERTIFICATE OF ANALYSIS · LOT RELEASE

CJC-1295 (no DAC) / Ipamorelin Blend 10 mg

Lot **PL-2609-024** · Lyophilized co-blend powder · Label claim **10 mg/vial**

TESTED IN-HOUSE BY PRYMALAB
In-house analytical result, not an independent third-party test

LOT / BATCH NUMBER	DATE MADE	DATE OF ANALYSIS	DATE ISSUED
PL-2609-024	2026-09-17	2026-09-18	2026-10-03
RETEST / EXPIRY	**UNITS IN LOT / SAMPLED**	**STORAGE**	**CERTIFICATE TYPE**
2028-09-17	250 / 2	−20 °C, desiccated, dark	Lot COA, 2 pages

COMPOUND	CAS NUMBER	MOLECULAR FORMULA	THEORETICAL MASS (AVG.)
CJC-1295 (no DAC) / Ipamorelin Blend Modified GRF(1–29) + Ipamorelin co-lyophilized 1:1 blend	863288-34-0 · 170851-70-4	$C_{152}H_{252}N_{44}O_{42}$ · $C_{38}H_{49}N_9O_5$	3367.9 · 711.9 Da (CJC-1295 (no DAC) / Ipamorelin)

COMPOSITION AND SEQUENCES
CJC-1295 (no DAC) 5 mg: H-Y-[D-Ala]-DAIFTQSYRKVLAQLSARKLLQDILSR-NH_2 (29 residues; C-terminal amide) · **Ipamorelin 5 mg**: H-Aib-H-[D-2-Nal]-[D-Phe]-K-NH_2 (5 residues; C-terminal amide)

PURITY · RP-HPLC, AREA %
99.69%
Specification 99.0% (blend, sum of both component peaks)

IDENTITY · OBSERVED MASS
2 / 2
Components confirmed by LC-MS within ± 1.0 Da: 3368.1 / 712.1 Da

OVERALL CONCLUSION
PASS
Meets every specification for the tests performed

TEST	METHOD	SPECIFICATION	RESULT	CONCLUSION
Appearance	Visual inspection	White to off-white lyophilized powder	White lyophilized powder	PASS
Identity (mass)	LC-MS, ESI positive	Theoretical ± 1.0 Da, each component	3368.1 Da (CJC-1295 No DAC) · 712.1 Da (Ipamorelin)	PASS
Purity	RP-HPLC, UV 214 nm, area%	99.0% (blend)	99.69%	PASS
Largest impurity	RP-HPLC, area%	1.0%	0.18% at 7.05 min	PASS
Water content	Karl Fischer	5.0%	1.7%	PASS
Bacterial endotoxin	LAL, kinetic	≤ 5.0 EU/mg	0.05 EU/mg	PASS
Sterility / bioburden	USP <71>, membrane filtration	No growth after 14 days	No growth; conforms	PASS

1 HPLC CHROMATOGRAM

Data file: CJI_PL-2609-024_20260918_01 · Column C18 150 × 4.6 mm · UV 214 nm



Baseline uncropped; both component peaks and impurities labeled with retention times. Peak table on page 2.

For laboratory research use only. Not for human or veterinary use.

2 HPLC METHOD · CJI_PL-2609-024_20260918_01

INSTRUMENT Waters Alliance e2695, SN E2695B742	DATA SYSTEM Empower 3, Build 3471	DETECTOR Waters 2998 PDA
DETECTION WAVELENGTH 214 nm	COLUMN C18, 150 × 4.6 mm, 3.5 µm	COLUMN PART / SERIAL 186003326 / USHX18211
MOBILE PHASE A 0.1% TFA in water	MOBILE PHASE B 0.1% TFA in MeCN	FLOW RATE 1.0 mL/min
COLUMN TEMPERATURE 30 °C	INJECTION VOLUME 20 µL	RUN TIME 10 min
SAMPLE DILUENT Water, 0.1% TFA	SAMPLE CONCENTRATION 0.5 mg/mL (blend)	BLANK INJECTED Yes

REFERENCE STANDARD
In-house CJC/Ipamorelin blend reference, lot PL-RS-CJI-01, component RTs 5.00 / 6.36 min

3 GRADIENT PROGRAM

TIME (MIN)	% A	% B
0.0	95	5
3.0	95	5
7.0	55	45
8.0	20	80
9.0	20	80
10.0	95	5

4 PEAK TABLE

PEAK	RT (MIN)	AREA	AREA %
1	4.18	5,590	0.13
2 (Ipamorelin)	5.02	2,196,010	51.07
3 (CJC-1295 No DAC)	6.38	2,090,660	48.62
4	7.05	7,740	0.18
Total		4,300,000	100.00

Threshold 0.05% area · Blank subtracted: Y · Purity = main peak area ÷ total integrated area × 100, to two decimals; main peak shaded.

5 IDENTITY BY MASS SPECTROMETRY

INSTRUMENT Waters Acquity SQD	IONIZATION / MODE ESI, positive	OBSERVED M/Z (CHARGE) 712.8 · 734.8 (1+ [M+H]/[M+Na], Ipamorelin) · 1123.7 (3+ [M+3H]) · 843.0 (4+ [M+4H], CJC-1295 No DAC)
DECONVOLUTED MASS 3368.1 / 712.1 Da (by component)	THEORETICAL MASS 3367.9 / 711.9 Da (CJC-1295 No DAC / Ipamorelin)	DIFFERENCE ≤ +0.2 Da (per component)
DATA FILE CJI_PL-2609-024_MS_20260918_01	ACCEPTANCE Within ± 1.0 Da (each component)	

6 ADDITIONAL TESTS

TEST	PERFORMED BY	DATE	REPORT / FILE
Water content	PrymaLab	2026-09-19	RPT-KF-PL-2609-024
Endotoxin	PrymaLab	2026-09-20	RPT-ET-PL-2609-024
Sterility	PrymaLab	2026-10-02	RPT-ST-PL-2609-024
Peptide content (blend assay)	PrymaLab	2026-09-19	RPT-AA-PL-2609-024

A test with no row here was not performed on this lot. Outside-laboratory reports, when used, are attached and named below.

7 REVIEW AND APPROVAL

D. Reynolds
Analyst · performed the testing
Date: 2026-09-18

S. Calloway
Reviewer · checked against raw data
Date: 2026-10-02

M. Phelps
Lab Director · approved for release
Date: 2026-10-03

Raw data retained at: Empower 3 (Build 3471), project PL-2609, archive PL-ARC-2026; 7 years. Outside laboratory reports attached: None

This certificate applies only to the lot named above and describes the units sampled from it; it is not a guarantee for any other lot. Results were obtained by PrymaLab LLC's own laboratory (PrymaLab Analytical) and are not an independent third-party test unless an outside report is attached. Reference values (CAS, formula, theoretical mass, sequence) are literature values for the named compound. Authenticity: scan the QR code or enter certificate number PL-26-00024 and key GFEZEMLN at prymalab.net/testing/verify; any copy that does not match the record shown there is not a PrymaLab certificate. PrymaLab LLC products are supplied for laboratory research use only and not for human or veterinary use.